



Greenlam LabGuardian Chemical Spot Test – Test Protocol: SEFA PL 2010-Clause 8.1

S.NO.	CHEMICAL	TEST METHOD	RATINGS	TYPICAL VALUES – LABGUARDIAN
1	Acetate, Amyl	A	0-1	0
2	Acetate, Ethyl	A	0-2	0
3	Acetic Acid 98%	B	0-3	0
4	Acetone	A	0-1	0
5	Acid Dichromate 5%	B	0-1	0
6	Alcohol, Butyl	A	0-1	0
7	Alcohol, Ethyl	A	0	0
8	Alcohol, Methyl	A	0-1	0
9	Ammonium Hydroxide 28%	B	0	0
10	Benzene	A	0-2	0
11	Carbon Tetrachloride	A	0-1	0
12	Chloroform	A	0-2	0
13	Chromic Acid 60%	B	0-2	0
14	Cresol	A	0-2	0
15	Dichloroacetic Acid	A	0-3	0
16	Dimethylformamide	A	0-2	0
17	Dioxane	A	0-2	0
18	Ethyl Ether	A	0-1	0
19	Formaldehyde 37%	A	0-1	0
20	Formic Acid 90%	B	0-3	0
21	Furfural	A	0-3	0
22	Gasoline	A	0	0
23	Hydrofluoric Acid 37%	B	0-2	0
24	Hydrofluoric Acid 48%	B	0-3	0
25	Hydrogen Peroxide 30%	B	0-1	0
26	Iodine, Tincture of	B	0-2	2
27	Methyl Ethyl Ketone	A	0-2	0
28	Methylene Chloride	A	0-2	0
29	Monochlorobenzene	A	0-2	0
30	Naphthalene	A	0-1	0
31	Nitric Acid 20%	B	0-1	0
32	Nitric Acid 30%	B	0-1	0
33	Nitric Acid 70%	B	0-3	0
34	Phenol 90%	A	0-2	0
35	Phosphoric Acid 85%	B	0-1	0
36	Silver Nitrate, Saturated	B	0	0
37	Sodium Hydroxide 10%	B	0	0
38	Sodium Hydroxide 20%	B	0	0
39	Sodium Hydroxide 40%	B	0-1	0
40	Sodium Hydroxide, Flake	B	0	0
41	Sodium Sulfide, Saturated	B	0	0
42	Sulfuric Acid 33%	B	0	0
43	Sulfuric Acid 77%	B	0	0
44	Sulfuric Acid 96%	B	02-03	2
45	Sulfuric Acid 77%, and Nitric Acid 70%, equal parts	B	02-03	1
46	Toluene	A	0-1	0
47	Trichloroethylene	A	0-1	0
48	Xylene	A	0-1	0
49	Zinc Chloride, Saturated	B	0	0